

# TrackSense® Pro X Wireless Data Logger

### 8 Bar Pressure Sensor

#### **Key Features & Benefits**

- ✓ Unique interchangeable sensor option
- ✓ Real-time data option with Sky module
- ✓ Works in conjunction with the ValSuite® validation & calibration software
- ✓ FDA 21 CFR Part 11 compliant
- ✓ Pressure Range of 20 mBar to 8 Bar Absolute
- ✓ Industry leading accuracy of ± 0.25% Full Scale
- Covers various applications within pharmaceutical, food and medical industries









## TrackSense® Pro X Data Loggers

The wireless TrackSense® Pro X data loggers are made of a high resistant stainless steel with cutting edge technology that allows for immensely accurate and stable measurements when performing various thermal processes.

All TrackSense® Pro X data loggers are configured with interchangeable sensors able to measure temperature, relative humidity, vacuum, pressure, conductivity and CO<sub>2</sub>.

Ellab's data loggers are easily activated and read by the Multi reader station. When combined with the Sky module, data loggers can provide real time data via wireless communication. Utilizing the numerous functions of the FDA 21 CFR, Part 11 compliant ValSuite™ software, data is easily analyzed and distributed through various report options.

Interested in this product? Contact sales today



# **Technical Specifications**

Sensor with this logger configuration:	8 Bar Pressure Sensor
Pressure Measuring Range:	20 mBar to 8 Bar ABS
Measuring Principle:	Piezoresistive
Sensor Element:	Strain Gauge
Accuracy:	± 0.25% Full Scale
Logger with this sensor configuration:	Pro X
Operating Pressure:	0.001 mBar to 10 Bar ABS
Operating Temperature:	0 to +150 °C
House Material:	316L Stainless Steel
Diameter:	25 mm
Length:	44 mm
Weight with Battery:	48 Grams
Memory Capacity:	60,000 Data Points / 30,000 Samples
Minimum Sample Rate:	0.33 Second
Maximum Sample Rate:	24 Hours
Maximum Start Delay:	14 Days
Intrinsically Safe:	Ex II1GD Ex ia IIC T3 Ga, -60°C ≤ Tamb ≤ +105°C
Time Accuracy:	± 5 Seconds Per 24 Hours
Battery:	TSP Standard Battery

If equipment is used in ATEX environment, special conditions for safe use are stated in ATEX certificates, section 17 must be considered.