

Operator: Validation Manager Vessel: Incubator
Process: Validation Product: Pharma
Session Start: 22-11-2013 08:30:02 Time Zone: UTC offset 01:00:00
Session Stop: 22-11-2013 08:30:24
Session Name: Incubator Validation 1
Session Text:



Limit Report

Name: Report 1
Description: Limit Report Number 1

Input parameters

Criteria: Passed when 23,00 °C < Value < 24,00 °C
Triggering : All, Temperature
Data Series: LA 01, LA 02, LC 03, LA 04, LA 05, LA 06, LA 08

Range

Time: 11-05-2010 15:14:21 to 12-05-2010 09:48:11

Triggering Interval

from	to
11-05-2010 15:59:41	11-05-2010 15:59:51
11-05-2010 16:00:41	11-05-2010 16:00:51
11-05-2010 16:01:11	11-05-2010 18:31:31
11-05-2010 18:32:21	11-05-2010 18:34:01
11-05-2010 18:34:31	11-05-2010 18:35:01
11-05-2010 18:35:21	11-05-2010 18:36:11
11-05-2010 18:37:11	11-05-2010 18:38:21
11-05-2010 18:39:41	11-05-2010 18:39:51
11-05-2010 18:40:41	11-05-2010 18:42:41
11-05-2010 18:43:11	11-05-2010 18:48:41
11-05-2010 18:49:01	11-05-2010 18:49:11
11-05-2010 18:50:01	11-05-2010 18:51:31
11-05-2010 18:52:01	11-05-2010 23:26:01
11-05-2010 23:26:21	11-05-2010 23:27:51
11-05-2010 23:28:11	11-05-2010 23:28:21
11-05-2010 23:28:41	11-05-2010 23:29:11
11-05-2010 23:29:31	11-05-2010 23:30:11
11-05-2010 23:31:11	11-05-2010 23:31:21
11-05-2010 23:32:41	11-05-2010 23:33:01
11-05-2010 23:33:31	11-05-2010 23:33:41
11-05-2010 23:34:41	11-05-2010 23:35:31
11-05-2010 23:40:11	11-05-2010 23:40:21
11-05-2010 23:42:41	11-05-2010 23:42:51
12-05-2010 09:42:51	12-05-2010 09:43:01

Summary Time: 07:23:00