

Operator: Validation Manager  
 Process: Validation  
 Session Start: 22-11-2013 08:30:02  
 Session Stop: 22-11-2013 08:30:24  
 Session Name: Incubator Validation 1  
 Session Text:

Vessel: Incubator  
 Product: Pharma  
 Time Zone: UTC offset 01:00:00



## Statistics

Name: Statistics

Description:

## Input parameters

Data Series Type: Temperature

Data series: LA 02, LA 01, LC 03, LA 04, LA 05, LA 06, LA 08

Start Time: 11-05-2010 15:15:41

Stop Time: 12-05-2010 09:43:01

Output: Global Data analysis

Functions: Min., Max., Average, Delta, Standard Deviation

## Global Data analysis

Min.	22,51 °C	12-05-2010 07:48:51	LA 02
		12-05-2010 07:50:21	LA 02
		12-05-2010 07:53:51	LA 02
		12-05-2010 08:06:01	LA 02
		12-05-2010 08:11:21	LA 02
		12-05-2010 08:11:41	LA 02
Max.	25,08 °C	11-05-2010 15:16:41	LA 01
		11-05-2010 15:16:51	LA 01
		11-05-2010 15:17:01	LA 01
Average	23,05 °C		
Delta	2,57 °C		
Standard Deviation	0,37 °C		

## Data Series analysis

Min average	22,99 °C	LC 03
Max average	23,15 °C	LA 08