

Operator: Validation Manager Vessel: Stability Chamb  
 Process: Validation Product: Pharma  
 Session Start: 22-11-2013 08:54:31 Time Zone: UTC offset 01:00:00  
 Session Stop: 22-11-2013 08:54:38  
 Session Name: Stability Chamber Validation  
 Session Text:



## Statistics

Name: Statistics on second run  
 Description: Heat distribution

## Input parameters

Data Series Type: Temperature  
 Data series: LA 01, LA 06, LC 02, LA 03, LA 04, LA 05, LA 08  
 Start Time: 11-05-2010 13:39:55  
 Stop Time: 11-05-2010 14:40:28  
 Output: Global Data analysis  
 Functions: Min., Max., Average, Delta, Standard Deviation

## Global Data analysis

Min.	22,65 °C	11-05-2010 14:19:04	LA 04
		11-05-2010 14:19:05	LA 04
		11-05-2010 14:19:06	LA 04
		11-05-2010 14:19:13	LA 04
		11-05-2010 14:19:14	LA 04
		11-05-2010 14:19:15	LA 04
		11-05-2010 14:19:16	LA 04
		11-05-2010 14:19:17	LA 04
Max.	26,57 °C	11-05-2010 13:40:01	LA 08
Average	23,85 °C		
Delta	3,92 °C		
Standard Deviation	0,99 °C		

## Data Series analysis

Min average	23,02 °C	LA 05
Max average	24,92 °C	LA 06