

Operator: Validation Manager Vessel: WD unit
 Process: Validation Product: Pharma Medical
 Session Start: 22-11-2013 09:09:28 Time Zone: UTC offset 01:00:00
 Session Stop: 22-11-2013 09:09:50
 Session Name: Washer Disinfectant Validation
 Session Text:



Statistics

Name: Statistics

Description:

Input parameters

Data Series Type: Temperature
 Data series: Top Load, Middle Load, Bottom Load, Hot Point, Bottom Cart, Top Cart, Middle Cart, Temp Sensor, Cold Point, Bowl
 Start Time: 02-12-2010 13:00:29
 Stop Time: 02-12-2010 13:05:45
 Output: Global Data analysis
 Functions: Min., Max., Average, Delta

Global Data analysis

Min.	55,52 °C	02-12-2010 13:05:45	Bottom Cart
Max.	56,37 °C	02-12-2010 13:00:30	Bowl
		02-12-2010 13:00:31	Bowl
		02-12-2010 13:00:32	Bowl
Average	55,87 °C		
Delta	0,85 °C		

Data Series analysis

Min average	55,82 °C	Bottom Cart
Max average	55,96 °C	Bowl